



[Optics](#) ➤ [Optical Specifications](#) ➤ [Cosmetic Surface Quality](#) ➤ [Cemented Surfaces](#)

[Home](#)

[Optics Guide Contents](#)

[Fundamental Optics](#) ➤

[Gaussian Beam Optics](#) ➤

[Optical Specifications](#) ➤

[Material Properties](#) ➤

[Optical Coatings](#) ➤

[Optics Glossary](#)

Cosmetic Surface Quality

Cemented Surfaces

Because a cemented interface is considered a lens surface, specified surface quality standards apply. Edge separation at a cemented interface cannot extend into the element more than half the distance to the element clear aperture up to a maximum of 1.0 mm. The sum of edge separations deeper than 0.5 mm cannot exceed 10% of the element perimeter.

[Previous](#) [Next](#)

Optics Guide © 2004 Melles Griot Inc.